

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081522\
 Data File : FF011261.D
 Signal(s) : FID2B.ch
 Acq On : 15 Aug 2022 20:28
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Aug 16 03:39:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF072822.M
 Quant Title :
 QLast Update : Thu Jul 28 14:46:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.799	5595198	51.480 ug/ml
Target Compounds			
2) N-DECANE	4.359	6697302	53.762 ug/ml
3) N-DODECANE	6.529	6721642	53.397 ug/ml
4) N-TETRADECANE	8.353	6688526	53.275 ug/ml
5) N-HEXADECANE	9.957	6685271	52.260 ug/ml
6) N-OCTADECANE	11.397	6675601	51.831 ug/ml
7) N-EICOSANE	12.705	6561289	51.323 ug/ml
8) N-DOCOSANE	13.901	6332546	50.768 ug/ml
10) N-TETRACOSANE	15.003	6296399	50.611 ug/ml
11) N-HEXACOSANE	16.022	6213668	50.781 ug/ml
12) N-OCTACOSANE	16.970	6224521	51.564 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081522\
Data File : FF011261.D
Signal(s) : FID2B.ch
Acq On : 15 Aug 2022 20:28
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Aug 16 03:39:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF072822.M
Quant Title :
QLast Update : Thu Jul 28 14:46:22 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

