

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081619\
 Data File : FF005299.D
 Signal(s) : FID2B.ch
 Acq On : 16 Aug 2019 20:06
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Aug 16 22:50:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
 Quant Title :
 QLast Update : Tue Jul 30 14:33:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.657	5460423	49.582 ug/ml
Target Compounds			
2) N-DECANE	4.373	7030535	48.799 ug/ml
3) N-DODECANE	6.396	6977291	47.819 ug/ml
4) N-TETRADECANE	8.190	6766233	47.009 ug/ml
5) N-HEXADECANE	9.792	6633814	46.787 ug/ml
6) N-OCTADECANE	11.237	6407460	46.796 ug/ml
7) N-EICOSANE	12.550	6347309	47.307 ug/ml
8) N-DOCOSANE	13.753	6254129	48.340 ug/ml
10) N-TETRACOSANE	14.861	6269603	49.764 ug/ml
11) N-HEXACOSANE	15.888	6302567	51.007 ug/ml
12) N-OCTACOSANE	16.842	6431743	51.490 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF081619\
Data File : FF005299.D
Signal(s) : FID2B.ch
Acq On : 16 Aug 2019 20:06
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Aug 16 22:50:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
Quant Title :
QLast Update : Tue Jul 30 14:33:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

