

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082118\
 Data File : FF001909.D
 Signal(s) : FID2B.ch
 Acq On : 21 Aug 2018 21:58
 Operator : AJ\SJ
 Sample : J4573-01 5X
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Aug 22 01:29:33 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
 Quant Title :
 QLast Update : Fri Aug 03 12:00:27 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.303	412418	3.265 ug/ml
Target Compounds			
7) N-EICOSANE	12.281f	509776	3.420 ug/ml
10) N-TETRACOSANE	14.509	341659	2.395 ug/ml
12) N-OCTACOSANE	16.420f	425305	3.033 ug/ml
16) N-HEXATRIACONTANE	19.704	57132	0.513 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082118\
Data File : FF001909.D
Signal(s) : FID2B.ch
Acq On : 21 Aug 2018 21:58
Operator : AJ\SJ
Sample : J4573-01 5X
Misc :
ALS Vial : 67 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: autoint1.e
Quant Time: Aug 22 01:29:33 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073018.M
Quant Title :
QLast Update : Fri Aug 03 12:00:27 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#7 N-EICOSANE


R.T.: 12.281 min
Delta R.T.: 0.052 min
Response: 509776
Conc: 3.42 ug/ml

Instrument :
FID_F
ClientSampleId :




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.303 min
Delta R.T.: -0.016 min
Response: 412418
Conc: 3.27 ug/ml




#10 N-TETRACOSANE

R.T.: 14.509 min
Delta R.T.: -0.012 min
Response: 341659
Conc: 2.40 ug/ml



#12 N-OCTACOSANE

R.T.: 16.420 min
Delta R.T.: -0.063 min
Response: 425305
Conc: 3.03 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 19.704 min
Delta R.T.: -0.012 min
Response: 57132
Conc: 0.51 ug/ml

Instrument :
FID_F
ClientSampleId :