

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082219\
 Data File : FF005315.D
 Signal(s) : FID2B.ch
 Acq On : 22 Aug 2019 18:40
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 23 00:44:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
 Quant Title :
 QLast Update : Tue Jul 30 14:33:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.647	5496811	49.913 ug/ml
Target Compounds			
2) N-DECANE	4.367	7312469	50.756 ug/ml
3) N-DODECANE	6.389	7220815	49.488 ug/ml
4) N-TETRADECANE	8.182	6961482	48.366 ug/ml
5) N-HEXADECANE	9.785	6774355	47.778 ug/ml
6) N-OCTADECANE	11.227	6495006	47.435 ug/ml
7) N-EICOSANE	12.540	6422916	47.870 ug/ml
8) N-DOCOSANE	13.743	6325147	48.889 ug/ml
10) N-TETRACOSANE	14.852	6348885	50.393 ug/ml
11) N-HEXACOSANE	15.877	6392964	51.739 ug/ml
12) N-OCTACOSANE	16.831	6452585	51.656 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082219\
Data File : FF005315.D
Signal(s) : FID2B.ch
Acq On : 22 Aug 2019 18:40
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 23 00:44:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
Quant Title :
QLast Update : Tue Jul 30 14:33:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

