

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082525\
 Data File : FF016196.D
 Signal(s) : FID2B.ch
 Acq On : 25 Aug 2025 11:20
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Aug 26 04:20:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF080625.M
 Quant Title :
 QLast Update : Wed Aug 06 13:50:48 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.084	5083548	45.543 ug/ml
Target Compounds			
2) N-DECANE	4.600	5316615	45.666 ug/ml
3) N-DODECANE	6.779	5403923	45.745 ug/ml
4) N-TETRADECANE	8.613	5510186	45.575 ug/ml
5) N-HEXADECANE	10.223	5543344	45.609 ug/ml
6) N-OCTADECANE	11.669	5701777	45.556 ug/ml
7) N-EICOSANE	12.982	5875092	45.404 ug/ml
8) N-DOCOSANE	14.182	5733150	45.348 ug/ml
10) N-TETRACOSANE	15.288	5744702	45.327 ug/ml
11) N-HEXACOSANE	16.312	5587037	45.133 ug/ml
12) N-OCTACOSANE	17.265	5608813	44.812 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082525\
Data File : FF016196.D
Signal(s) : FID2B.ch
Acq On : 25 Aug 2025 11:20
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Aug 26 04:20:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF080625.M
Quant Title :
QLast Update : Wed Aug 06 13:50:48 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

