

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082719\
 Data File : FF005360.D
 Signal(s) : FID2B.ch
 Acq On : 27 Aug 2019 15:14
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Aug 27 22:33:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
 Quant Title :
 QLast Update : Tue Jul 30 14:33:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.642	5870032	53.302 ug/ml
Target Compounds			
2) N-DECANE	4.364	7522089	52.211 ug/ml
3) N-DODECANE	6.386	7553494	51.768 ug/ml
4) N-TETRADECANE	8.180	7376272	51.248 ug/ml
5) N-HEXADECANE	9.781	7253702	51.159 ug/ml
6) N-OCTADECANE	11.225	7001669	51.135 ug/ml
7) N-EICOSANE	12.539	6896810	51.402 ug/ml
8) N-DOCOSANE	13.740	6749031	52.166 ug/ml
10) N-TETRACOSANE	14.848	6720771	53.345 ug/ml
11) N-HEXACOSANE	15.873	6735770	54.513 ug/ml
12) N-OCTACOSANE	16.828	6820413	54.601 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082719\
Data File : FF005360.D
Signal(s) : FID2B.ch
Acq On : 27 Aug 2019 15:14
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Aug 27 22:33:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF073019.M
Quant Title :
QLast Update : Tue Jul 30 14:33:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

