

rteres

Instrument :

FID_F

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 08/28/2025

Supervised By :mohammad ahmed 08/29/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF08272
Data File : FF016242.D
Signal(s) : FID2B.ch
Acq On : 27 Aug 2025 14:10
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF080625.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.597	4.562	4.679	BB	628878	6503732	90.20%	8.690%
2	6.776	6.737	6.847	BB	646486	6613747	91.72%	8.837%
3	8.610	8.576	8.689	BB	654883	6743216	93.52%	9.010%
4	10.220	10.182	10.294	BB	623283	6792587	94.20%	9.076%
5	11.666	11.619	11.729	BB	596075	6989658	96.94%	9.340%
6	12.979	12.936	13.052	BB	612588	7210654	100.00%	9.635%
7	14.180	14.141	14.236	VB	562497	7034826	97.56%	9.400%
8	15.080	14.982	15.151	BV	451819	6223660	86.31%	8.316%
9	15.285	15.224	15.337	BB	548269	7031539	97.52%	9.396%
10	16.309	16.252	16.361	BB	512879	6831242	94.74%	9.128%
11	17.262	17.201	17.319	BB	508354	6863423	95.18%	9.171%
Sum of corrected areas:						74838284		

FF080625.M Thu Aug 28 03:55:13 2025