

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082923AL\
 Data File : FF013450.D
 Signal(s) : FID2B.ch
 Acq On : 30 Aug 2023 08:46
 Operator : YP\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Aug 30 11:04:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH 081023.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 10 16:05:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.890	2789432	20.302 ug/ml
Spiked Amount 50.000		Recovery =	40.60%
12) S 1-chlorooctadecane (S...	13.329	2104193	20.079 ug/ml
Spiked Amount 50.000		Recovery =	40.16%
Target Compounds			
1) T n-Nonane (C9)	3.273	2255815	21.112 ug/ml
2) T n-Decane (C10)	4.527	2292628	21.393 ug/ml
3) T A~Naphthalene (C11.7)	6.229	2531117	21.380 ug/ml
4) T n-Dodecane (C12)	6.695	2323407	21.472 ug/ml
5) T A~2-methylnaphthalene...	7.325	2485915	20.972 ug/ml
6) T n-Tetradecane (C14)	8.523	2312421	21.422 ug/ml
7) T n-Hexadecane (C16)	10.130	2367127	21.837 ug/ml
8) T n-Octadecane (C18)	11.575	2376694	21.705 ug/ml
10) T n-Eicosane (C20)	12.885	2314217	21.090 ug/ml
11) T n-Heneicosane (C21)	13.498	2288234	20.827 ug/ml
13) T n-Docosane (C22)	14.084	2261615	20.600 ug/ml
14) T n-Tetracosane (C24)	15.187	2227772	20.277 ug/ml
15) T n-Hexacosane (C26)	16.209	2164662	19.806 ug/ml
16) T n-Octacosane (C28)	17.159	2103769	19.484 ug/ml
17) T n-Tricontane (C30)	18.046	2114169	19.332 ug/ml
18) T n-Dotriacontane (C32)	18.879	2102165	18.610 ug/ml
19) T n-Tetratriacontane (C34)	19.663	2062008	19.412 ug/ml
20) T n-Hexatriacontane (C36)	20.403	2129734	20.206 ug/ml
21) T n-Octatriacontane (C38)	21.132	2119625	20.986 ug/ml
22) T n-Tetracontane (C40)	22.029	2291096	23.051 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF082923AL\
Data File : FF013450.D
Signal(s) : FID2B.ch
Acq On : 30 Aug 2023 08:46
Operator : YP\AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
Quant Time: Aug 30 11:04:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\Aliphatic EPH 081023.M
Quant Title : GC Extractables
QLast Update : Thu Aug 10 16:05:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

