

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF083018\
 Data File : FF001974.D
 Signal(s) : FID2B.ch
 Acq On : 30 Aug 2018 13:26
 Operator : AJ\SJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 30 14:19:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 13:37:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.262	1278303	10.647 ug/ml
Target Compounds			
1) N-OCTANE	2.018	1111018	9.055 ug/ml
2) N-DECANE	4.068	1092535	8.853 ug/ml
3) N-DODECANE	6.071	1208281	9.371 ug/ml
4) N-TETRADECANE	7.853	1307131	9.867 ug/ml
5) N-HEXADECANE	9.442	1368391	10.144 ug/ml
6) N-OCTADECANE	10.875	1409994	10.309 ug/ml
7) N-EICOSANE	12.176	1444556	10.530 ug/ml
8) N-DOCOSANE	13.367	1451574	10.633 ug/ml
10) N-TETRACOSANE	14.464	1449035	10.689 ug/ml
11) N-HEXACOSANE	15.481	1453691	10.725 ug/ml
12) N-OCTACOSANE	16.426	1472841	10.846 ug/ml
13) N-TRIACONTANE	17.310	1493651	10.972 ug/ml
14) N-DOTRIACONTANE	18.139	1451635	10.997 ug/ml
15) N-TETRATRIACONTANE	18.920	1366960	10.705 ug/ml
16) N-HEXATRIACONTANE	19.658	1250590	10.428 ug/ml
17) N-OCTATRIACONTANE	20.357	1090731	10.299 ug/ml
18) N-TETRACONTANE	21.122	1049695	10.231 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF083018\
Data File : FF001974.D
Signal(s) : FID2B.ch
Acq On : 30 Aug 2018 13:26
Operator : AJ\SJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 30 14:19:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 13:37:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

