

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF083018\
 Data File : FF001976.D
 Signal(s) : FID2B.ch
 Acq On : 30 Aug 2018 14:26
 Operator : AJ\SJ
 Sample : FF083018ICV
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 30 14:48:21 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.268	5955750	47.518 ug/ml
Target Compounds			
1) N-OCTANE	2.016	6133387	51.989 ug/ml
2) N-DECANE	4.068	6152725	53.089 ug/ml
3) N-DODECANE	6.073	6416959	51.283 ug/ml
4) N-TETRADECANE	7.854	6603435	49.836 ug/ml
5) N-HEXADECANE	9.445	6724889	49.103 ug/ml
6) N-OCTADECANE	10.878	6811197	48.571 ug/ml
7) N-EICOSANE	12.180	6799001	47.700 ug/ml
8) N-DOCOSANE	13.372	6783309	47.611 ug/ml
10) N-TETRACOSANE	14.470	6718549	47.344 ug/ml
11) N-HEXACOSANE	15.488	6706483	47.152 ug/ml
12) N-OCTACOSANE	16.433	6705290	46.738 ug/ml
13) N-TRIACONTANE	17.316	6709868	46.273 ug/ml
14) N-DOTRIACONTANE	18.147	6515950	46.218 ug/ml
15) N-TETRATRIACONTANE	18.928	6336443	47.119 ug/ml
16) N-HEXATRIACONTANE	19.666	5952690	47.797 ug/ml
17) N-OCTATRIACONTANE	20.366	5245810	47.906 ug/ml
18) N-TETRACONTANE	21.134	5052158	47.719 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF083018\
Data File : FF001976.D
Signal(s) : FID2B.ch
Acq On : 30 Aug 2018 14:26
Operator : AJ\SJ
Sample : FF083018ICV
Misc :
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 30 14:48:21 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

