

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090120\
 Data File : FF007439.D
 Signal(s) : FID2B.ch
 Acq On : 01 Sep 2020 19:19
 Operator : DD\AJ
 Sample : FF090120ICV
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 FF090120ICV

Manual Integrations
 APPROVED

Ankita
 9/2/2020 3:45:15 PM

Integration File: Sample.e
 Quant Time: Sep 02 11:22:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
 Quant Title :
 QLast Update : Wed Sep 02 11:20:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.743	7189905	53.043 ug/ml
Target Compounds			
1) N-OCTANE	2.345	7712436	53.325 ug/ml
2) N-DECANE	4.456	7953712	53.353 ug/ml
3) N-DODECANE	6.484	8115028	53.446 ug/ml
4) N-TETRADECANE	8.281	8164863	53.453 ug/ml
5) N-HEXADECANE	9.884	8359968	53.355 ug/ml
6) N-OCTADECANE	11.328	8402189	53.268 ug/ml
7) N-EICOSANE	12.641	8263353	53.152 ug/ml
8) N-DOCOSANE	13.842	8081211	53.088 ug/ml
10) N-TETRACOSANE	14.948	7957074	52.978 ug/ml
11) N-HEXACOSANE	15.971	7663700	52.846 ug/ml
12) N-OCTACOSANE	16.923	7366334	52.641 ug/ml
13) N-TRIACONTANE	17.812	7038778	52.558 ug/ml
14) N-DOTRIACONTANE	18.648	6534072	52.616 ug/ml
15) N-TETRATRIACONTANE	19.433	6055732	52.312 ug/ml
16) N-HEXATRIACONTANE	20.176	5553122	50.877 ug/ml
17) N-OCTATRIACONTANE	20.934	5039214	52.126 ug/ml
18) N-TETRACONTANE	21.880	5021522	49.229 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090120\
Data File : FF007439.D
Signal(s) : FID2B.ch
Acq On : 01 Sep 2020 19:19
Operator : DD\AJ
Sample : FF090120ICV
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
FID_F
ClientSampled :
FF090120ICV

Manual Integrations
APPROVED

Ankita
9/2/2020 3:45:15 PM

Integration File: Sample.e
Quant Time: Sep 02 11:22:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:20:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

