

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090320\
 Data File : FF007458.D
 Signal(s) : FID2B.ch
 Acq On : 03 Sep 2020 13:43
 Operator : DD\AJ
 Sample : PB131360BSD
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 04 00:40:19 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
 Quant Title :
 QLast Update : Wed Sep 02 11:29:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.739	3218918	21.886 ug/ml
Target Compounds			
1) N-OCTANE	2.346	3547501	22.896 ug/ml
2) N-DECANE	4.454	3679641	23.069 ug/ml
3) N-DODECANE	6.480	3743389	23.010 ug/ml
4) N-TETRADECANE	8.277	3740806	22.733 ug/ml
5) N-HEXADECANE	9.880	3795106	22.488 ug/ml
6) N-OCTADECANE	11.323	3789983	22.241 ug/ml
7) N-EICOSANE	12.635	3708023	22.055 ug/ml
8) N-DOCOSANE	13.836	3613096	21.930 ug/ml
10) N-TETRACOSANE	14.942	3543223	21.725 ug/ml
11) N-HEXACOSANE	15.966	3416174	21.611 ug/ml
12) N-OCTACOSANE	16.918	3311921	21.581 ug/ml
13) N-TRIACONTANE	17.808	3205528	21.688 ug/ml
14) N-DOTRIACONTANE	18.643	3020994	21.772 ug/ml
15) N-TETRATRIACONTANE	19.429	2720223	21.555 ug/ml
16) N-HEXATRIACONTANE	20.171	2474273	20.705 ug/ml
17) N-OCTATRIACONTANE	20.929	2279473	21.347 ug/ml
18) N-TETRACONTANE	21.873	2566227	23.130 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090320\
Data File : FF007458.D
Signal(s) : FID2B.ch
Acq On : 03 Sep 2020 13:43
Operator : DD\AJ
Sample : PB131360BSD
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 04 00:40:19 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:29:30 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

