

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090320\
 Data File : FF007460.D
 Signal(s) : FID2B.ch
 Acq On : 03 Sep 2020 14:45
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 04 00:40:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
 Quant Title :
 QLast Update : Wed Sep 02 11:29:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.742	7889256	53.641 ug/ml
Target Compounds			
1) N-OCTANE	2.344	8808055	56.848 ug/ml
2) N-DECANE	4.455	9141817	57.314 ug/ml
3) N-DODECANE	6.482	9277313	57.027 ug/ml
4) N-TETRADECANE	8.279	9252575	56.228 ug/ml
5) N-HEXADECANE	9.882	9387306	55.626 ug/ml
6) N-OCTADECANE	11.327	9360225	54.929 ug/ml
7) N-EICOSANE	12.639	9140699	54.368 ug/ml
8) N-DOCOSANE	13.840	8882958	53.915 ug/ml
10) N-TETRACOSANE	14.947	8700116	53.343 ug/ml
11) N-HEXACOSANE	15.970	8356003	52.860 ug/ml
12) N-OCTACOSANE	16.922	8014037	52.220 ug/ml
13) N-TRIACONTANE	17.813	7647443	51.741 ug/ml
14) N-DOTRIACONTANE	18.647	7060903	50.887 ug/ml
15) N-TETRATRIACONTANE	19.433	6512557	51.605 ug/ml
16) N-HEXATRIACONTANE	20.174	5922749	49.561 ug/ml
17) N-OCTATRIACONTANE	20.935	5533104	51.816 ug/ml
18) N-TETRACONTANE	21.878	5418619	48.840 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090320\
Data File : FF007460.D
Signal(s) : FID2B.ch
Acq On : 03 Sep 2020 14:45
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 04 00:40:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:29:30 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

