

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090419\
 Data File : FF005418.D
 Signal(s) : FID2B.ch
 Acq On : 04 Sep 2019 15:25
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 04 15:47:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090319.M
 Quant Title :
 QLast Update : Wed Sep 04 03:16:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.632	5903737	55.171 ug/ml
Target Compounds			
1) N-OCTANE	2.265	7244956	50.358 ug/ml
2) N-DECANE	4.357	7378158	51.605 ug/ml
3) N-DODECANE	6.377	7455343	52.009 ug/ml
4) N-TETRADECANE	8.170	7433222	52.721 ug/ml
5) N-HEXADECANE	9.771	7437356	53.176 ug/ml
6) N-OCTADECANE	11.215	7257359	53.940 ug/ml
7) N-EICOSANE	12.528	7156846	54.572 ug/ml
8) N-DOCOSANE	13.730	6914303	54.916 ug/ml
10) N-TETRACOSANE	14.837	6684147	54.655 ug/ml
11) N-HEXACOSANE	15.863	6500014	54.301 ug/ml
12) N-OCTACOSANE	16.817	6420725	53.809 ug/ml
13) N-TRIACONTANE	17.710	6449692	53.318 ug/ml
14) N-DOTRIACONTANE	18.549	6442629	51.973 ug/ml
15) N-TETRATRIACONTANE	19.340	6401461	53.121 ug/ml
16) N-HEXATRIACONTANE	20.088	6134568	51.111 ug/mlm
17) N-OCTATRIACONTANE	20.832	5753270	52.732 ug/ml
18) N-TETRACONTANE	21.739	6149231	48.367 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090419\
Data File : FF005418.D
Signal(s) : FID2B.ch
Acq On : 04 Sep 2019 15:25
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 04 15:47:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090319.M
Quant Title :
QLast Update : Wed Sep 04 03:16:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

