

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090619\
 Data File : FF005443.D
 Signal(s) : FID2B.ch
 Acq On : 06 Sep 2019 08:59
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Sep 06 23:37:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090319.M
 Quant Title :
 QLast Update : Wed Sep 04 03:16:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.634	5318769	49.704 ug/ml
Target Compounds			
2) N-DECANE	4.358	7394900	51.722 ug/ml
3) N-DODECANE	6.378	7283044	50.807 ug/ml
4) N-TETRADECANE	8.171	7055801	50.044 ug/ml
5) N-HEXADECANE	9.772	6909644	49.403 ug/ml
6) N-OCTADECANE	11.216	6630234	49.279 ug/ml
7) N-EICOSANE	12.528	6451315	49.192 ug/ml
8) N-DOCOSANE	13.730	6188443	49.151 ug/ml
10) N-TETRACOSANE	14.838	6051438	49.482 ug/ml
11) N-HEXACOSANE	15.863	6057246	50.602 ug/ml
12) N-OCTACOSANE	16.819	6240493	52.298 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090619\
Data File : FF005443.D
Signal(s) : FID2B.ch
Acq On : 06 Sep 2019 08:59
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Sep 06 23:37:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090319.M
Quant Title :
QLast Update : Wed Sep 04 03:16:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

