

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090718\
 Data File : FF002051.D
 Signal(s) : FID2B.ch
 Acq On : 08 Sep 2018 01:37
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 08 02:16:07 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.279	6244730	49.823 ug/ml
Target Compounds			
1) N-OCTANE	2.013	6577592	55.755 ug/ml
2) N-DECANE	4.068	6591242	56.873 ug/ml
3) N-DODECANE	6.076	6800560	54.349 ug/ml
4) N-TETRADECANE	7.859	6958936	52.519 ug/ml
5) N-HEXADECANE	9.451	7054289	51.508 ug/ml
6) N-OCTADECANE	10.886	7113235	50.724 ug/ml
7) N-EICOSANE	12.190	7082481	49.689 ug/ml
8) N-DOCOSANE	13.383	7066584	49.600 ug/ml
10) N-TETRACOSANE	14.482	7034126	49.567 ug/ml
11) N-HEXACOSANE	15.499	6974678	49.037 ug/ml
12) N-OCTACOSANE	16.446	6956319	48.487 ug/ml
13) N-TRIACONTANE	17.331	6741281	46.490 ug/ml
14) N-DOTRIACONTANE	18.161	5713823	40.528 ug/ml
15) N-TETRATRIACONTANE	18.942	4193903	31.187 ug/mlm
16) N-HEXATRIACONTANE	19.681	3170883	25.460 ug/ml
17) N-OCTATRIACONTANE	20.381	2477181	22.622 ug/ml
18) N-TETRACONTANE	21.153	2710103	25.598 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF090718\
Data File : FF002051.D
Signal(s) : FID2B.ch
Acq On : 08 Sep 2018 01:37
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 08 02:16:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

