

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091018\
Data File : FF002069.D
Signal(s) : FID2B.ch
Acq On : 10 Sep 2018 14:09
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 11 02:31:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds

3)	N-DODECANE	6.073	4432126	35.421 ug/ml
12)	N-OCTACOSANE	16.441	4664196	32.511 ug/ml
13)	N-TRIACONTANE	17.324	279348	1.926 ug/ml
14)	N-DOTRIACONTANE	18.155	377501	2.678 ug/ml
15)	N-TETRATRIACONTANE	18.936	283632	2.109 ug/ml
16)	N-HEXATRIACONTANE	19.676	125755	1.010 ug/ml
17)	N-OCTATRIACONTANE	20.376	130605	1.193 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091018\
Data File : FF002069.D
Signal(s) : FID2B.ch
Acq On : 10 Sep 2018 14:09
Operator : AJ\SJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 11 02:31:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.073 min
Delta R.T.: 0.002 min
Response: 4432126
Conc: 35.42 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T.: 14.266 min
Response: 0
Conc: N.D.




#12 N-OCTACOSANE

R.T.: 16.441 min
Delta R.T.: 0.011 min
Response: 4664196
Conc: 32.51 ug/ml




#13 N-TRIACONTANE

R.T.: 17.324 min
Delta R.T.: 0.010 min
Response: 279348
Conc: 1.93 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.155 min
Delta R.T.: 0.011 min
Response: 377501
Conc: 2.68 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 18.936 min
Delta R.T.: 0.011 min
Response: 283632
Conc: 2.11 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 19.676 min
Delta R.T.: 0.012 min
Response: 125755
Conc: 1.01 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.376 min
Delta R.T.: 0.013 min
Response: 130605
Conc: 1.19 ug/ml