

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091018\
 Data File : FF002068.D
 Signal(s) : FID2B.ch
 Acq On : 10 Sep 2018 09:40
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 11 02:31:02 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.277	5654010	45.110 ug/ml
Target Compounds			
1) N-OCTANE	2.013	6613474	56.059 ug/ml
2) N-DECANE	4.066	6684158	57.675 ug/ml
3) N-DODECANE	6.073	6816364	54.475 ug/ml
4) N-TETRADECANE	7.857	6917725	52.208 ug/ml
5) N-HEXADECANE	9.449	6918772	50.519 ug/ml
6) N-OCTADECANE	10.884	6902122	49.219 ug/ml
7) N-EICOSANE	12.187	6778827	47.559 ug/ml
8) N-DOCOSANE	13.381	6694347	46.987 ug/ml
10) N-TETRACOSANE	14.480	6590348	46.440 ug/ml
11) N-HEXACOSANE	15.497	6495815	45.671 ug/ml
12) N-OCTACOSANE	16.445	6468892	45.090 ug/ml
13) N-TRIACONTANE	17.329	6290790	43.383 ug/ml
14) N-DOTRIACONTANE	18.159	5317575	37.718 ug/ml
15) N-TETRATRIACONTANE	18.940	3934872	29.261 ug/mlm
16) N-HEXATRIACONTANE	19.680	2886869	23.180 ug/mlm
17) N-OCTATRIACONTANE	20.378	2384821	21.779 ug/ml
18) N-TETRACONTANE	21.150	2174041	20.534 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091018\
Data File : FF002068.D
Signal(s) : FID2B.ch
Acq On : 10 Sep 2018 09:40
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 11 02:31:02 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

