

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091018\
 Data File : FF002081.D
 Signal(s) : FID2B.ch
 Acq On : 10 Sep 2018 20:08
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 11 02:32:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.275	5938046	47.376 ug/ml
Target Compounds			
1) N-OCTANE	2.013	6763642	57.332 ug/ml
2) N-DECANE	4.066	6816869	58.820 ug/ml
3) N-DODECANE	6.073	6986851	55.838 ug/ml
4) N-TETRADECANE	7.857	7099481	53.580 ug/ml
5) N-HEXADECANE	9.448	7141415	52.144 ug/ml
6) N-OCTADECANE	10.883	7139167	50.909 ug/ml
7) N-EICOSANE	12.186	7030416	49.324 ug/ml
8) N-DOCOSANE	13.379	6976436	48.967 ug/ml
10) N-TETRACOSANE	14.478	6898417	48.611 ug/ml
11) N-HEXACOSANE	15.496	6833171	48.043 ug/ml
12) N-OCTACOSANE	16.442	6850686	47.751 ug/ml
13) N-TRIACONTANE	17.327	6729460	46.408 ug/ml
14) N-DOTRIACONTANE	18.156	5832649	41.371 ug/ml
15) N-TETRATRIACONTANE	18.937	4567929	33.968 ug/ml
16) N-HEXATRIACONTANE	19.677	3585116	28.786 ug/ml
17) N-OCTATRIACONTANE	20.378	2918160	26.649 ug/ml
18) N-TETRACONTANE	21.147	2726424	25.752 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091018\
Data File : FF002081.D
Signal(s) : FID2B.ch
Acq On : 10 Sep 2018 20:08
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 11 02:32:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

