

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091020\
 Data File : FF007492.D
 Signal(s) : FID2B.ch
 Acq On : 10 Sep 2020 10:54
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 11 03:35:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
 Quant Title :
 QLast Update : Wed Sep 02 11:29:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.737	6149620	41.813 ug/ml
Target Compounds			
2) N-DECANE	4.452	7357690	46.129 ug/ml
3) N-DODECANE	6.479	7461232	45.863 ug/ml
4) N-TETRADECANE	8.276	7415535	45.065 ug/ml
5) N-HEXADECANE	9.878	7525545	44.594 ug/ml
6) N-OCTADECANE	11.323	7489068	43.949 ug/ml
7) N-EICOSANE	12.635	7277954	43.288 ug/ml
8) N-DOCOSANE	13.836	7008046	42.535 ug/ml
10) N-TETRACOSANE	14.941	6768410	41.499 ug/ml
11) N-HEXACOSANE	15.964	6389770	40.421 ug/ml
12) N-OCTACOSANE	16.917	6039284	39.352 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091020\
Data File : FF007492.D
Signal(s) : FID2B.ch
Acq On : 10 Sep 2020 10:54
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 11 03:35:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:29:30 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

