

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091118\
 Data File : FF002100.D
 Signal(s) : FID2B.ch
 Acq On : 11 Sep 2018 13:15
 Operator : AJ\SJ
 Sample : PB112751BS
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 12 04:23:43 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.271	2699786	21.540 ug/ml
Target Compounds			
1) N-OCTANE	2.017	1824950	15.469 ug/ml
2) N-DECANE	4.062	2600274	22.437 ug/ml
3) N-DODECANE	6.069	2993484	23.923 ug/ml
4) N-TETRADECANE	7.852	3272745	24.700 ug/ml
5) N-HEXADECANE	9.444	3421802	24.985 ug/ml
6) N-OCTADECANE	10.878	3478619	24.806 ug/ml
7) N-EICOSANE	12.182	3454824	24.238 ug/ml
8) N-DOCOSANE	13.375	3406809	23.912 ug/ml
10) N-TETRACOSANE	14.475	3354620	23.639 ug/ml
11) N-HEXACOSANE	15.491	3196861	22.476 ug/ml
12) N-OCTACOSANE	16.438	3089442	21.534 ug/ml
13) N-TRIACONTANE	17.323	2953340	20.367 ug/ml
14) N-DOTRIACONTANE	18.153	2887184	20.479 ug/ml
15) N-TETRATRIACONTANE	18.934	2716931	20.204 ug/ml
16) N-HEXATRIACONTANE	19.675	2537415	20.374 ug/ml
17) N-OCTATRIACONTANE	20.377	2412147	22.028 ug/ml
18) N-TETRACONTANE	21.145	2581372	24.382 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091118\
Data File : FF002100.D
Signal(s) : FID2B.ch
Acq On : 11 Sep 2018 13:15
Operator : AJ\SJ
Sample : PB112751BS
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 12 04:23:43 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

