

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091118\
 Data File : FF002106.D
 Signal(s) : FID2B.ch
 Acq On : 11 Sep 2018 16:46
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 12 04:46:36 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.275	6778566	54.083 ug/ml
Target Compounds			
1) N-OCTANE	2.009	7425149	62.939 ug/ml
2) N-DECANE	4.064	7593254	65.519 ug/ml
3) N-DODECANE	6.071	7816584	62.469 ug/ml
4) N-TETRADECANE	7.856	7986833	60.277 ug/ml
5) N-HEXADECANE	9.447	8044350	58.737 ug/ml
6) N-OCTADECANE	10.882	8070179	57.548 ug/ml
7) N-EICOSANE	12.186	7931551	55.646 ug/ml
8) N-DOCOSANE	13.379	7810053	54.818 ug/ml
10) N-TETRACOSANE	14.478	7753758	54.639 ug/ml
11) N-HEXACOSANE	15.496	7589056	53.357 ug/ml
12) N-OCTACOSANE	16.443	7531681	52.498 ug/ml
13) N-TRIACONTANE	17.328	7345232	50.655 ug/ml
14) N-DOTRIACONTANE	18.157	6343819	44.997 ug/ml
15) N-TETRATRIACONTANE	18.938	4899117	36.431 ug/ml
16) N-HEXATRIACONTANE	19.678	3893122	31.259 ug/ml
17) N-OCTATRIACONTANE	20.378	3087472	28.195 ug/ml
18) N-TETRACONTANE	21.147	2978623	28.134 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091118\
Data File : FF002106.D
Signal(s) : FID2B.ch
Acq On : 11 Sep 2018 16:46
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 12 04:46:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

