

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091120\
 Data File : FF007535.D
 Signal(s) : FID2B.ch
 Acq On : 11 Sep 2020 23:01
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 12 00:14:47 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
 Quant Title :
 QLast Update : Wed Sep 02 11:29:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.739	6587964	44.793 ug/ml
Target Compounds			
1) N-OCTANE	2.342	7473682	48.236 ug/ml
2) N-DECANE	4.454	8569754	53.727 ug/ml
3) N-DODECANE	6.481	8595406	52.835 ug/ml
4) N-TETRADECANE	8.278	8431684	51.240 ug/ml
5) N-HEXADECANE	9.880	8440249	50.014 ug/ml
6) N-OCTADECANE	11.325	8282516	48.605 ug/ml
7) N-EICOSANE	12.636	7931897	47.178 ug/ml
8) N-DOCOSANE	13.838	7534802	45.732 ug/ml
10) N-TETRACOSANE	14.943	7222286	44.282 ug/ml
11) N-HEXACOSANE	15.966	6851620	43.343 ug/ml
12) N-OCTACOSANE	16.918	6639915	43.266 ug/ml
13) N-TRIACONTANE	17.808	6628954	44.850 ug/ml
14) N-DOTRIACONTANE	18.642	6493472	46.798 ug/ml
15) N-TETRATRIACONTANE	19.427	6358782	50.387 ug/ml
16) N-HEXATRIACONTANE	20.169	6057461	50.688 ug/ml
17) N-OCTATRIACONTANE	20.925	5769726	54.032 ug/ml
18) N-TETRACONTANE	21.870	5771706	52.023 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091120\
Data File : FF007535.D
Signal(s) : FID2B.ch
Acq On : 11 Sep 2020 23:01
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 12 00:14:47 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:29:30 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

