

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091319\
 Data File : FF005546.D
 Signal(s) : FID2B.ch
 Acq On : 13 Sep 2019 08:58
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 14 04:24:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090319.M
 Quant Title :
 QLast Update : Wed Sep 04 03:16:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.622	5446730	50.900 ug/ml
Target Compounds			
1) N-OCTANE	2.262	7435942	51.685 ug/ml
2) N-DECANE	4.351	7386440	51.663 ug/ml
3) N-DODECANE	6.371	7240520	50.510 ug/ml
4) N-TETRADECANE	8.163	6974661	49.468 ug/ml
5) N-HEXADECANE	9.764	6897780	49.318 ug/ml
6) N-OCTADECANE	11.207	6636523	49.326 ug/ml
7) N-EICOSANE	12.519	6489420	49.483 ug/ml
8) N-DOCOSANE	13.720	6283245	49.904 ug/ml
10) N-TETRACOSANE	14.827	6218186	50.845 ug/ml
11) N-HEXACOSANE	15.852	6252585	52.234 ug/ml
12) N-OCTACOSANE	16.806	6373672	53.414 ug/ml
13) N-TRIACONTANE	17.699	6466979	53.461 ug/ml
14) N-DOTRIACONTANE	18.536	6511077	52.525 ug/ml
15) N-TETRATRIACONTANE	19.327	6298699	52.268 ug/ml
16) N-HEXATRIACONTANE	20.072	5822178	48.508 ug/mlm
17) N-OCTATRIACONTANE	20.817	5399007	49.485 ug/ml
18) N-TETRACONTANE	21.719	5492501	43.201 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091319\
Data File : FF005546.D
Signal(s) : FID2B.ch
Acq On : 13 Sep 2019 08:58
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 14 04:24:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090319.M
Quant Title :
QLast Update : Wed Sep 04 03:16:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

