

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091420\
 Data File : FF007540.D
 Signal(s) : FID2B.ch
 Acq On : 14 Sep 2020 12:03
 Operator : DD\AJ
 Sample : PB131545BS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 14 15:28:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
 Quant Title :
 QLast Update : Wed Sep 02 11:29:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.736	2420475	16.457 ug/ml
Target Compounds			
1) N-OCTANE	2.344	2722320	17.570 ug/ml
2) N-DECANE	4.452	3258262	20.427 ug/ml
3) N-DODECANE	6.477	3289745	20.222 ug/ml
4) N-TETRADECANE	8.274	3271234	19.879 ug/ml
5) N-HEXADECANE	9.877	3207953	19.009 ug/ml
6) N-OCTADECANE	11.322	3175443	18.635 ug/ml
7) N-EICOSANE	12.634	3057405	18.185 ug/ml
8) N-DOCOSANE	13.834	2919187	17.718 ug/ml
10) N-TETRACOSANE	14.940	2787500	17.091 ug/ml
11) N-HEXACOSANE	15.963	2636072	16.676 ug/ml
12) N-OCTACOSANE	16.916	2551227	16.624 ug/ml
13) N-TRIACONTANE	17.805	2497761	16.899 ug/ml
14) N-DOTRIACONTANE	18.641	2359557	17.005 ug/ml
15) N-TETRATRIACONTANE	19.428	2054818	16.282 ug/ml
16) N-HEXATRIACONTANE	20.172	1708300	14.295 ug/ml
17) N-OCTATRIACONTANE	20.929	1408373	13.189 ug/mlm
18) N-TETRACONTANE	21.874	1388499	12.515 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091420\
Data File : FF007540.D
Signal(s) : FID2B.ch
Acq On : 14 Sep 2020 12:03
Operator : DD\AJ
Sample : PB131545BS
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 14 15:28:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:29:30 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

