

rteres

Area Percent Report

Instrument :
FID_F
LabSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Ankita
9/18/2020 10:10:07 AM

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF091720\
Data File : FF007589.D
Signal(s) : FID2B.ch
Acq On : 17 Sep 2020 13:41
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Title :

Signal : FID2B.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.334	2.301	2.418	BB	1064084	7844728	98.07%	6.500%
2	4.447	4.408	4.521	BB	884892	7999086	100.00%	6.628%
3	6.474	6.435	6.555	BB	803925	7954205	99.44%	6.590%
4	8.273	8.195	8.338	BB	733702	7743362	96.80%	6.416%
5	9.877	9.838	9.941	BB	689456	7691512	96.15%	6.373%
6	11.321	11.271	11.381	BB	660669	7513836	93.93%	6.226%
7	12.634	12.595	12.698	BB	618515	7179973	89.76%	5.949%
8	13.836	13.761	13.898	BB	554008	6830967	85.40%	5.660%
9	14.738	14.691	14.811	BB	436848	5982294	74.79%	4.957%
10	14.942	14.901	15.001	BB	515451	6599684	82.51%	5.468%
11	15.965	15.811	16.021	BB	491307	6341342	79.28%	5.254%
12	16.919	16.848	16.975	BB	473435	6203423	77.55%	5.140%
13	17.808	17.741	17.868	BB	453506	6180082	77.26%	5.120%
14	18.645	18.571	18.695	BB	410959	6050781	75.64%	5.013%
15	19.430	19.358	19.478	BV	419684	5961119	74.52%	4.939%
16	20.174	20.127	20.229	M	386077	5699411	71.25%	4.722%
17	20.933	20.874	20.992	M	295199	5436129	67.96%	4.504%
18	21.879	21.781	22.000	BBA	224435	5481486	68.53%	4.542%
Sum of corrected areas:						120693419		

FF090120.M Fri Sep 18 09:05:47 2020