

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092018\
 Data File : FF002156.D
 Signal(s) : FID2B.ch
 Acq On : 20 Sep 2018 10:37
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 21 00:48:06 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.269	5533300	44.147 ug/ml
Target Compounds			
1) N-OCTANE	2.010	6088073	51.605 ug/ml
2) N-DECANE	4.059	5898387	50.895 ug/ml
3) N-DODECANE	6.066	6148921	49.141 ug/ml
4) N-TETRADECANE	7.850	6290587	47.475 ug/ml
5) N-HEXADECANE	9.442	6402358	46.748 ug/ml
6) N-OCTADECANE	10.876	6440939	45.930 ug/ml
7) N-EICOSANE	12.179	6376619	44.737 ug/ml
8) N-DOCOSANE	13.372	6317938	44.345 ug/ml
10) N-TETRACOSANE	14.471	6246247	44.016 ug/ml
11) N-HEXACOSANE	15.487	6168246	43.368 ug/ml
12) N-OCTACOSANE	16.435	6152441	42.884 ug/ml
13) N-TRIACONTANE	17.320	6189841	42.687 ug/ml
14) N-DOTRIACONTANE	18.150	6184389	43.866 ug/ml
15) N-TETRATRIACONTANE	18.933	6294999	46.811 ug/ml
16) N-HEXATRIACONTANE	19.671	6170991	49.549 ug/ml
17) N-OCTATRIACONTANE	20.372	5613803	51.266 ug/ml
18) N-TETRACONTANE	21.141	5491859	51.872 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092018\
Data File : FF002156.D
Signal(s) : FID2B.ch
Acq On : 20 Sep 2018 10:37
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 21 00:48:06 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

