

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092018\
 Data File : FF002169.D
 Signal(s) : FID2B.ch
 Acq On : 20 Sep 2018 17:33
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 21 00:50:36 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.269	5443931	43.434 ug/ml
Target Compounds			
1) N-OCTANE	2.011	6050158	51.284 ug/ml
2) N-DECANE	4.060	5950920	51.348 ug/ml
3) N-DODECANE	6.067	6181504	49.401 ug/ml
4) N-TETRADECANE	7.851	6309106	47.615 ug/ml
5) N-HEXADECANE	9.442	6385537	46.625 ug/ml
6) N-OCTADECANE	10.876	6386491	45.542 ug/ml
7) N-EICOSANE	12.180	6298870	44.191 ug/ml
8) N-DOCOSANE	13.374	6227904	43.713 ug/ml
10) N-TETRACOSANE	14.472	6155226	43.374 ug/ml
11) N-HEXACOSANE	15.490	6096885	42.866 ug/ml
12) N-OCTACOSANE	16.436	6115657	42.628 ug/ml
13) N-TRIACONTANE	17.321	6219055	42.888 ug/ml
14) N-DOTRIACONTANE	18.151	6257812	44.387 ug/ml
15) N-TETRATRIACONTANE	18.934	6383904	47.472 ug/ml
16) N-HEXATRIACONTANE	19.674	6240971	50.111 ug/mlm
17) N-OCTATRIACONTANE	20.377	5745637	52.470 ug/ml
18) N-TETRACONTANE	21.145	5671370	53.567 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092018\
Data File : FF002169.D
Signal(s) : FID2B.ch
Acq On : 20 Sep 2018 17:33
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 21 00:50:36 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

