

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092118\
 Data File : FF002184.D
 Signal(s) : FID2B.ch
 Acq On : 21 Sep 2018 18:14
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Sep 22 00:30:08 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:20:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.268	6043516	48.218 ug/ml
Target Compounds			
2) N-DECANE	4.059	6455002	55.698 ug/ml
3) N-DODECANE	6.066	6703125	53.570 ug/ml
4) N-TETRADECANE	7.850	6884900	51.961 ug/ml
5) N-HEXADECANE	9.442	7011324	51.194 ug/ml
6) N-OCTADECANE	10.876	7063480	50.370 ug/ml
7) N-EICOSANE	12.180	6998430	49.099 ug/ml
8) N-DOCOSANE	13.372	6906819	48.478 ug/ml
10) N-TETRACOSANE	14.471	6800741	47.923 ug/ml
11) N-HEXACOSANE	15.488	6685024	47.001 ug/ml
12) N-OCTACOSANE	16.436	6634988	46.247 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092118\
Data File : FF002184.D
Signal(s) : FID2B.ch
Acq On : 21 Sep 2018 18:14
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Sep 22 00:30:08 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF083018.M
Quant Title :
QLast Update : Thu Aug 30 14:20:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

