

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092420\
Data File : FF007655.D
Signal(s) : FID2B.ch
Acq On : 24 Sep 2020 11:35
Operator : DD\AJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 25 02:10:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:29:30 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds

3)	N-DODECANE	6.474	4148630	25.501 ug/ml
12)	N-OCTACOSANE	16.911	3449835	22.479 ug/ml
13)	N-TRIACONTANE	17.799	264273	1.788 ug/ml
14)	N-DOTRIACONTANE	18.634	305381	2.201 ug/ml
15)	N-TETRATRIACONTANE	19.421	229031	1.815 ug/ml
16)	N-HEXATRIACONTANE	20.165	161712	1.353 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF092420\
Data File : FF007655.D
Signal(s) : FID2B.ch
Acq On : 24 Sep 2020 11:35
Operator : DD\AJ
Sample : RT MARKER
Misc :
ALS Vial : 54 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 25 02:10:48 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF090120.M
Quant Title :
QLast Update : Wed Sep 02 11:29:30 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#3 N-DODECANE

R.T.: 6.474 min
Delta R.T.: -0.009 min
Response: 4148630
Conc: 25.50 ug/ml




#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T. : 14.743 min
Response: 0
Conc: N.D.




#12 N-OCTACOSANE

R.T.: 16.911 min
Delta R.T.: -0.012 min
Response: 3449835
Conc: 22.48 ug/ml




#13 N-TRIACONTANE

R.T.: 17.799 min
Delta R.T.: -0.014 min
Response: 264273
Conc: 1.79 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.634 min
Delta R.T.: -0.013 min
Response: 305381
Conc: 2.20 ug/ml



#15 N-TETRATRIACONTANE

R.T.: 19.421 min
Delta R.T.: -0.012 min
Response: 229031
Conc: 1.81 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.165 min
Delta R.T.: -0.012 min
Response: 161712
Conc: 1.35 ug/ml