

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100118\
 Data File : FF002281.D
 Signal(s) : FID2B.ch
 Acq On : 01 Oct 2018 13:10
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Oct 02 03:45:23 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:14:42 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.640	6750325	47.330 ug/ml
Target Compounds			
2) N-DECANE	4.392	7683626	47.433 ug/ml
3) N-DODECANE	6.410	7804223	47.441 ug/ml
4) N-TETRADECANE	8.201	7888371	47.356 ug/ml
5) N-HEXADECANE	9.799	7940374	47.413 ug/ml
6) N-OCTADECANE	11.239	7959021	47.401 ug/ml
7) N-EICOSANE	12.546	7844179	47.364 ug/ml
8) N-DOCOSANE	13.744	7718525	47.303 ug/ml
10) N-TETRACOSANE	14.845	7544660	47.184 ug/ml
11) N-HEXACOSANE	15.864	7395831	47.220 ug/ml
12) N-OCTACOSANE	16.811	7218029	47.186 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100118\
Data File : FF002281.D
Signal(s) : FID2B.ch
Acq On : 01 Oct 2018 13:10
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Oct 02 03:45:23 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
Quant Title :
QLast Update : Mon Oct 01 12:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

