

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100118\
 Data File : FF002284.D
 Signal(s) : FID2B.ch
 Acq On : 01 Oct 2018 14:45
 Operator : AJ\SJ
 Sample : PB113351BSD
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 PB113351BSD

Integration File: Sample.e
 Quant Time: Oct 02 03:46:00 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:14:42 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.638	2747898	19.267 ug/ml
Target Compounds			
2) N-DECANE	4.390	3091464	19.085 ug/ml
3) N-DODECANE	6.408	3145875	19.124 ug/ml
4) N-TETRADECANE	8.197	3192152	19.163 ug/ml
5) N-HEXADECANE	9.795	3217959	19.215 ug/ml
6) N-OCTADECANE	11.235	3230607	19.240 ug/ml
7) N-EICOSANE	12.543	3191964	19.273 ug/ml
8) N-DOCOSANE	13.739	3138975	19.237 ug/ml
10) N-TETRACOSANE	14.840	3071739	19.210 ug/ml
11) N-HEXACOSANE	15.860	3014667	19.248 ug/ml
12) N-OCTACOSANE	16.807	2963583	19.374 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100118\
Data File : FF002284.D
Signal(s) : FID2B.ch
Acq On : 01 Oct 2018 14:45
Operator : AJ\SJ
Sample : PB113351BSD
Misc :
ALS Vial : 56 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
PB113351BSD

Integration File: Sample.e
Quant Time: Oct 02 03:46:00 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
Quant Title :
QLast Update : Mon Oct 01 12:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

