

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100518\
 Data File : FF002351.D
 Signal(s) : FID2B.ch
 Acq On : 05 Oct 2018 14:58
 Operator : AJ\SJ
 Sample : J5199-01MS 2X
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 05 23:48:09 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:14:42 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.656	1110203	7.784 ug/ml
Target Compounds			
1) N-OCTANE	2.287	678302	4.291 ug/ml
2) N-DECANE	4.389	1155006	7.130 ug/ml
3) N-DODECANE	6.410	1245283	7.570 ug/mlm
4) N-TETRADECANE	8.203	1366572	8.204 ug/ml
5) N-HEXADECANE	9.804	1568921	9.368 ug/ml
6) N-OCTADECANE	11.246	1438604	8.568 ug/ml
7) N-EICOSANE	12.556	1534322	9.264 ug/ml
8) N-DOCOSANE	13.756	1417374	8.686 ug/ml
10) N-TETRACOSANE	14.861	1185505	7.414 ug/ml
11) N-HEXACOSANE	15.882	1212195	7.739 ug/mlm
12) N-OCTACOSANE	16.831	1257950	8.224 ug/mlm
13) N-TRIACONTANE	17.720	1226379	8.199 ug/mlm
14) N-DOTRIACONTANE	18.557	1253229	8.696 ug/ml
15) N-TETRATRIACONTANE	19.342	1174309	8.455 ug/ml
16) N-HEXATRIACONTANE	20.044	15606	0.121 ug/mlm
17) N-OCTATRIACONTANE	20.828	2129982	18.648 ug/ml
18) N-TETRACONTANE	21.741f	1103572	9.764 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100518\
Data File : FF002351.D
Signal(s) : FID2B.ch
Acq On : 05 Oct 2018 14:58
Operator : AJ\SJ
Sample : J5199-01MS 2X
Misc :
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 05 23:48:09 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
Quant Title :
QLast Update : Mon Oct 01 12:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

