

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100818\
 Data File : FF002364.D
 Signal(s) : FID2B.ch
 Acq On : 08 Oct 2018 11:28
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 09 02:07:18 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
 Quant Title :
 QLast Update : Mon Oct 01 12:14:42 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.661	5864159	41.116 ug/ml
Target Compounds			
1) N-OCTANE	2.284	7228680	45.731 ug/ml
2) N-DECANE	4.390	7378575	45.550 ug/ml
3) N-DODECANE	6.414	7423634	45.128 ug/ml
4) N-TETRADECANE	8.207	7415332	44.516 ug/ml
5) N-HEXADECANE	9.808	7363570	43.969 ug/ml
6) N-OCTADECANE	11.251	7274679	43.325 ug/ml
7) N-EICOSANE	12.561	7048190	42.558 ug/ml
8) N-DOCOSANE	13.761	6796051	41.650 ug/ml
10) N-TETRACOSANE	14.865	6512132	40.726 ug/ml
11) N-HEXACOSANE	15.886	6279984	40.095 ug/ml
12) N-OCTACOSANE	16.837	6095272	39.846 ug/ml
13) N-TRIACONTANE	17.726	6004075	40.140 ug/ml
14) N-DOTRIACONTANE	18.560	5882378	40.815 ug/ml
15) N-TETRATRIACONTANE	19.344	5893416	42.434 ug/ml
16) N-HEXATRIACONTANE	20.087	5789396	44.933 ug/ml
17) N-OCTATRIACONTANE	20.829	5242957	45.901 ug/ml
18) N-TETRACONTANE	21.742f	5204745	46.050 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100818\
Data File : FF002364.D
Signal(s) : FID2B.ch
Acq On : 08 Oct 2018 11:28
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 02:07:18 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100118.M
Quant Title :
QLast Update : Mon Oct 01 12:14:42 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

