

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100919\
 Data File : FF005756.D
 Signal(s) : FID2B.ch
 Acq On : 09 Oct 2019 10:11
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 09 14:36:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
 Quant Title :
 QLast Update : Thu Oct 03 04:07:51 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.676	6612110	50.854 ug/ml
Target Compounds			
2) N-DECANE	4.419	8295977	53.426 ug/ml
3) N-DODECANE	6.439	8369258	53.144 ug/ml
4) N-TETRADECANE	8.232	8286122	52.724 ug/ml
5) N-HEXADECANE	9.831	8264773	52.430 ug/ml
6) N-OCTADECANE	11.272	8067837	52.000 ug/ml
7) N-EICOSANE	12.580	7973283	51.497 ug/ml
8) N-DOCOSANE	13.777	7731085	50.941 ug/ml
10) N-TETRACOSANE	14.879	7517949	50.524 ug/ml
11) N-HEXACOSANE	15.899	7299801	50.036 ug/ml
12) N-OCTACOSANE	16.847	7080629	49.717 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF100919\
Data File : FF005756.D
Signal(s) : FID2B.ch
Acq On : 09 Oct 2019 10:11
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 09 14:36:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
Quant Title :
QLast Update : Thu Oct 03 04:07:51 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

