

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101019\
 Data File : FF005778.D
 Signal(s) : FID2B.ch
 Acq On : 10 Oct 2019 18:36
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 11 00:59:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
 Quant Title :
 QLast Update : Thu Oct 03 04:07:51 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.676	6365061	48.954 ug/ml
Target Compounds			
2) N-DECANE	4.422	8397103	54.077 ug/ml
3) N-DODECANE	6.441	8445943	53.631 ug/ml
4) N-TETRADECANE	8.234	8301402	52.822 ug/ml
5) N-HEXADECANE	9.833	8205857	52.057 ug/ml
6) N-OCTADECANE	11.273	7932457	51.128 ug/ml
7) N-EICOSANE	12.581	7772584	50.200 ug/ml
8) N-DOCOSANE	13.778	7483103	49.307 ug/ml
10) N-TETRACOSANE	14.880	7244392	48.686 ug/ml
11) N-HEXACOSANE	15.899	7041194	48.263 ug/ml
12) N-OCTACOSANE	16.847	6895544	48.417 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101019\
Data File : FF005778.D
Signal(s) : FID2B.ch
Acq On : 10 Oct 2019 18:36
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 11 00:59:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
Quant Title :
QLast Update : Thu Oct 03 04:07:51 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

