

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101024\
 Data File : FF014624.D
 Signal(s) : FID2B.ch
 Acq On : 10 Oct 2024 16:39
 Operator : YP\AJ
 Sample : P4368-10
 Misc : 2.5 PPM WATER LOD
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 LOD-MDL-WATER-04-QT4-2024

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 10/11/2024
 Supervised By :Ankita Jodhani 10/11/2024

Integration File: autoint1.e
 Quant Time: Oct 11 01:26:30 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100324.M
 Quant Title :
 QLast Update : Thu Oct 03 12:23:07 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.883	2267923	17.342 ug/ml
Target Compounds			
1) N-OCTANE	1.941	297290	2.223 ug/ml
2) N-DECANE	4.444	295785	2.182 ug/ml
3) N-DODECANE	6.610	288688	2.116 ug/ml
4) N-TETRADECANE	8.437	298714	2.170 ug/ml
5) N-HEXADECANE	10.044	297523	2.176 ug/ml
6) N-OCTADECANE	11.483	344196	2.404 ug/ml
7) N-EICOSANE	12.790	365533	2.461 ug/ml
8) N-DOCOSANE	13.985	380442	2.616 ug/ml
10) N-TETRACOSANE	15.087	387172	2.644 ug/ml
11) N-HEXACOSANE	16.109	364322	2.500 ug/ml
12) N-OCTACOSANE	17.058	355790	2.466 ug/ml
13) N-TRIACONTANE	17.945	363124	2.487 ug/ml
14) N-DOTRIACONTANE	18.778	304840	2.100 ug/ml
15) N-TETRATRIACONTANE	19.564	195865	1.421 ug/ml
16) N-HEXATRIACONTANE	20.336	848466	6.373 ug/mlm
17) N-OCTATRIACONTANE	21.025	110115	0.983 ug/mlm
18) N-TETRACONTANE	21.925f	92191	0.987 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101024\
Data File : FF014624.D
Signal(s) : FID2B.ch
Acq On : 10 Oct 2024 16:39
Operator : YP\AJ
Sample : P4368-10
Misc : 2.5 PPM WATER LOD
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
LOD-MDL-WATER-04-QT4-2024

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 10/11/2024
Supervised By :Ankita Jodhani 10/11/2024

Integration File: autoint1.e
Quant Time: Oct 11 01:26:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100324.M
Quant Title :
QLast Update : Thu Oct 03 12:23:07 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

