

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101419\
 Data File : FF005794.D
 Signal(s) : FID2B.ch
 Acq On : 14 Oct 2019 15:16
 Operator : DD\AJ
 Sample : K5286-01MSD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 15 00:38:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
 Quant Title :
 QLast Update : Thu Oct 03 04:07:51 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.674	1857563	14.287 ug/mlm
Target Compounds			
1) N-OCTANE	2.320	1778830	11.805 ug/ml
2) N-DECANE	4.419	2532317	16.308 ug/ml
3) N-DODECANE	6.437	2676958	16.998 ug/ml
4) N-TETRADECANE	8.230	2588752	16.472 ug/ml
5) N-HEXADECANE	9.829	2677445	16.985 ug/ml
6) N-OCTADECANE	11.270	2385051	15.373 ug/ml
7) N-EICOSANE	12.578	2600115	16.793 ug/ml
8) N-DOCOSANE	13.775	2415982	15.919 ug/ml
10) N-TETRACOSANE	14.878	2093591	14.070 ug/ml
11) N-HEXACOSANE	15.897	2241262	15.362 ug/ml
12) N-OCTACOSANE	16.847	2480201	17.415 ug/ml
13) N-TRIACONTANE	17.734	2313672	16.659 ug/ml
14) N-DOTRIACONTANE	18.566	3110325	23.212 ug/ml
15) N-TETRATRIACONTANE	19.353	2263790	18.720 ug/ml
16) N-HEXATRIACONTANE	20.093	1732731	16.044 ug/ml
17) N-OCTATRIACONTANE	20.828	1693725	17.113 ug/ml
18) N-TETRACONTANE	21.727	1670812	14.899 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101419\
Data File : FF005794.D
Signal(s) : FID2B.ch
Acq On : 14 Oct 2019 15:16
Operator : DD\AJ
Sample : K5286-01MSD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 15 00:38:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
Quant Title :
QLast Update : Thu Oct 03 04:07:51 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

