

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101519\
 Data File : FF005809.D
 Signal(s) : FID2B.ch
 Acq On : 15 Oct 2019 15:12
 Operator : DD\AJ
 Sample : K5355-03MSD
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 16 00:47:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
 Quant Title :
 QLast Update : Thu Oct 03 04:07:51 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.676	1854150	14.260 ug/ml
Target Compounds			
1) N-OCTANE	2.318	830688	5.513 ug/ml
2) N-DECANE	4.420	1621822	10.444 ug/ml
3) N-DODECANE	6.438	1971281	12.517 ug/ml
4) N-TETRADECANE	8.232	2112196	13.440 ug/ml
5) N-HEXADECANE	9.831	2176238	13.806 ug/ml
6) N-OCTADECANE	11.271	2054028	13.239 ug/mlm
7) N-EICOSANE	12.579	2112732	13.645 ug/ml
8) N-DOCOSANE	13.776	2147354	14.149 ug/ml
10) N-TETRACOSANE	14.878	1925708	12.942 ug/ml
11) N-HEXACOSANE	15.900	1935399	13.266 ug/ml
12) N-OCTACOSANE	16.849	2054877	14.428 ug/ml
13) N-TRIACONTANE	17.735	2222887	16.006 ug/ml
14) N-DOTRIACONTANE	18.568	2056568	15.348 ug/ml
15) N-TETRATRIACONTANE	19.353	1632176	13.497 ug/ml
16) N-HEXATRIACONTANE	20.096	1584712	14.674 ug/mlm
17) N-OCTATRIACONTANE	20.831	1879525	18.991 ug/ml
18) N-TETRACONTANE	21.733	1700841	15.167 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101519\
Data File : FF005809.D
Signal(s) : FID2B.ch
Acq On : 15 Oct 2019 15:12
Operator : DD\AJ
Sample : K5355-03MSD
Misc :
ALS Vial : 59 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 16 00:47:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
Quant Title :
QLast Update : Thu Oct 03 04:07:51 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

