

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101519\
 Data File : FF005820.D
 Signal(s) : FID2B.ch
 Acq On : 15 Oct 2019 20:38
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 16 00:49:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
 Quant Title :
 QLast Update : Thu Oct 03 04:07:51 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.675	5636514	43.350 ug/ml
Target Compounds			
1) N-OCTANE	2.317	6972782	46.274 ug/ml
2) N-DECANE	4.421	7482214	48.185 ug/ml
3) N-DODECANE	6.440	7472634	47.451 ug/ml
4) N-TETRADECANE	8.233	7299069	46.444 ug/ml
5) N-HEXADECANE	9.831	7172939	45.504 ug/ml
6) N-OCTADECANE	11.272	6898731	44.465 ug/ml
7) N-EICOSANE	12.580	6750582	43.600 ug/ml
8) N-DOCOSANE	13.777	6511828	42.907 ug/ml
10) N-TETRACOSANE	14.879	6350023	42.675 ug/ml
11) N-HEXACOSANE	15.899	6241533	42.782 ug/ml
12) N-OCTACOSANE	16.846	6213154	43.626 ug/ml
13) N-TRIACONTANE	17.733	6289024	45.283 ug/ml
14) N-DOTRIACONTANE	18.566	6178538	46.109 ug/ml
15) N-TETRATRIACONTANE	19.350	6054773	50.068 ug/mlm
16) N-HEXATRIACONTANE	20.091	5606968	51.918 ug/mlm
17) N-OCTATRIACONTANE	20.824	5304361	53.595 ug/ml
18) N-TETRACONTANE	21.726	5341690	47.633 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101519\
Data File : FF005820.D
Signal(s) : FID2B.ch
Acq On : 15 Oct 2019 20:38
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 16 00:49:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
Quant Title :
QLast Update : Thu Oct 03 04:07:51 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

