

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101520\
 Data File : FF007725.D
 Signal(s) : FID2B.ch
 Acq On : 15 Oct 2020 13:40
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 16 00:59:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100220.M
 Quant Title :
 QLast Update : Mon Oct 05 15:41:23 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.733	5470975	43.473 ug/ml
Target Compounds			
2) N-DECANE	4.447	6848829	48.285 ug/ml
3) N-DODECANE	6.473	6915174	47.857 ug/ml
4) N-TETRADECANE	8.270	6862898	47.072 ug/ml
5) N-HEXADECANE	9.874	6950469	46.490 ug/ml
6) N-OCTADECANE	11.318	6905599	45.836 ug/ml
7) N-EICOSANE	12.630	6701720	45.077 ug/ml
8) N-DOCOSANE	13.831	6440319	44.152 ug/ml
10) N-TETRACOSANE	14.937	6242277	43.273 ug/ml
11) N-HEXACOSANE	15.960	5946348	42.805 ug/ml
12) N-OCTACOSANE	16.912	5751101	42.939 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101520\
Data File : FF007725.D
Signal(s) : FID2B.ch
Acq On : 15 Oct 2020 13:40
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 16 00:59:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100220.M
Quant Title :
QLast Update : Mon Oct 05 15:41:23 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

