

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101521\
 Data File : FF009146.D
 Signal(s) : FID2B.ch
 Acq On : 15 Oct 2021 15:29
 Operator : AJ\MA
 Sample : M4068-03
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 MDL-SOIL-03-QT4-2021

Manual Integrations
 APPROVED

mohammad
 10/18/2021 4:38:22 PM

Integration File: autoint1.e
 Quant Time: Oct 16 00:16:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100721.M
 Quant Title :
 QLast Update : Fri Oct 08 02:25:33 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.709	2054038	18.732 ug/ml
Target Compounds			
2) N-DECANE	4.465	143870	1.135 ug/ml
3) N-DODECANE	6.482	85488	0.673 ug/ml
4) N-TETRADECANE	8.291	67905	0.542 ug/mlm
5) N-HEXADECANE	9.882	116294	0.916 ug/mlm
6) N-OCTADECANE	11.314	116908	0.907 ug/ml
7) N-EICOSANE	12.616	129568	1.006 ug/ml
8) N-DOCOSANE	13.812	129662	1.026 ug/ml
10) N-TETRACOSANE	14.914	124610	0.995 ug/mlm
11) N-HEXACOSANE	15.940	106481	0.880 ug/ml
12) N-OCTACOSANE	16.886	112184	0.958 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101521\
Data File : FF009146.D
Signal(s) : FID2B.ch
Acq On : 15 Oct 2021 15:29
Operator : AJ\MA
Sample : M4068-03
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
MDL-SOIL-03-QT4-2021

Manual Integrations
APPROVED

mohammad
10/18/2021 4:38:22 PM

Integration File: autoint1.e
Quant Time: Oct 16 00:16:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100721.M
Quant Title :
QLast Update : Fri Oct 08 02:25:33 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

