

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101718\
Data File : FF002423.D
Signal(s) : FID2B.ch
Acq On : 17 Oct 2018 15:45
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 17 16:39:25 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF101218.M
Quant Title :
QLast Update : Fri Oct 12 13:21:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.654	2923243	21.032 ug/ml
Target Compounds			
14) N-DOTRIACONTANE	18.554	142122	0.917 ug/ml
16) N-HEXATRIACONTANE	20.131	184849	1.348 ug/ml
18) N-TETRACONTANE	21.782	101870	0.855 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101718\
Data File : FF002423.D
Signal(s) : FID2B.ch
Acq On : 17 Oct 2018 15:45
Operator : AJ\SJ
Sample : I.BLK
Misc :
ALS Vial : 52 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 17 16:39:25 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF101218.M
Quant Title :
QLast Update : Fri Oct 12 13:21:40 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.654 min
Delta R.T.: -0.003 min
Response: 2923243
Conc: 21.03 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.554 min
Delta R.T.: -0.006 min
Response: 142122
Conc: 0.92 ug/ml



#16 N-HEXATRIACONTANE

R.T.: 20.131 min
Delta R.T.: 0.042 min
Response: 184849
Conc: 1.35 ug/ml



#18 N-TETRACONTANE

R.T.: 21.782 min
Delta R.T.: 0.036 min
Response: 101870
Conc: 0.86 ug/ml