

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101719\
 Data File : FF005836.D
 Signal(s) : FID2B.ch
 Acq On : 17 Oct 2019 20:25
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 18 00:45:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
 Quant Title :
 QLast Update : Thu Oct 03 04:07:51 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.671	6765726	52.035 ug/ml
Target Compounds			
1) N-OCTANE	2.318	8023471	53.246 ug/ml
2) N-DECANE	4.419	8772609	56.495 ug/ml
3) N-DODECANE	6.438	8853374	56.218 ug/ml
4) N-TETRADECANE	8.231	8745915	55.650 ug/ml
5) N-HEXADECANE	9.830	8686415	55.105 ug/ml
6) N-OCTADECANE	11.270	8417110	54.251 ug/ml
7) N-EICOSANE	12.578	8238061	53.207 ug/ml
8) N-DOCOSANE	13.773	7897990	52.041 ug/ml
10) N-TETRACOSANE	14.875	7586405	50.984 ug/ml
11) N-HEXACOSANE	15.895	7303952	50.064 ug/ml
12) N-OCTACOSANE	16.843	7134151	50.093 ug/ml
13) N-TRIACONTANE	17.730	7125932	51.309 ug/ml
14) N-DOTRIACONTANE	18.562	6909283	51.563 ug/ml
15) N-TETRATRIACONTANE	19.347	6679184	55.231 ug/mlm
16) N-HEXATRIACONTANE	20.087	6185822	57.278 ug/mlm
17) N-OCTATRIACONTANE	20.822	5765896	58.258 ug/ml
18) N-TETRACONTANE	21.722	5780739	51.548 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101719\
Data File : FF005836.D
Signal(s) : FID2B.ch
Acq On : 17 Oct 2019 20:25
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 18 00:45:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100219.M
Quant Title :
QLast Update : Thu Oct 03 04:07:51 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

