

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101821\
 Data File : FF009168.D
 Signal(s) : FID2B.ch
 Acq On : 18 Oct 2021 15:41
 Operator : AJ\MA
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 19 00:18:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100721.M
 Quant Title :
 QLast Update : Fri Oct 08 02:25:33 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.714	5371171	48.982 ug/ml
Target Compounds			
1) N-OCTANE	2.365	6509012	52.065 ug/ml
2) N-DECANE	4.467	6660696	52.552 ug/ml
3) N-DODECANE	6.485	6699358	52.734 ug/ml
4) N-TETRADECANE	8.278	6646619	53.010 ug/ml
5) N-HEXADECANE	9.875	6618321	52.129 ug/ml
6) N-OCTADECANE	11.314	6632540	51.450 ug/ml
7) N-EICOSANE	12.621	6509660	50.566 ug/ml
8) N-DOCOSANE	13.817	6260390	49.519 ug/ml
10) N-TETRACOSANE	14.918	6085250	48.613 ug/ml
11) N-HEXACOSANE	15.936	5794332	47.864 ug/ml
12) N-OCTACOSANE	16.884	5502695	47.006 ug/ml
13) N-TRIACONTANE	17.768	5208734	46.169 ug/ml
14) N-DOTRIACONTANE	18.598	4873772	46.152 ug/ml
15) N-TETRATRIACONTANE	19.379	4608263	47.153 ug/ml
16) N-HEXATRIACONTANE	20.117	4353330	48.965 ug/ml
17) N-OCTATRIACONTANE	20.856	4106246	50.296 ug/ml
18) N-TETRACONTANE	21.775	4057627	49.945 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101821\
Data File : FF009168.D
Signal(s) : FID2B.ch
Acq On : 18 Oct 2021 15:41
Operator : AJ\MA
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 19 00:18:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100721.M
Quant Title :
QLast Update : Fri Oct 08 02:25:33 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

