

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101920\
 Data File : FF007747.D
 Signal(s) : FID2B.ch
 Acq On : 19 Oct 2020 13:54
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Oct 19 14:07:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100220.M
 Quant Title :
 QLast Update : Mon Oct 05 15:41:23 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.730	5653591	44.924 ug/ml
Target Compounds			
1) N-OCTANE	2.336	7514998	54.042 ug/ml
2) N-DECANE	4.446	7618533	53.711 ug/ml
3) N-DODECANE	6.473	7590779	52.533 ug/ml
4) N-TETRADECANE	8.270	7439195	51.025 ug/ml
5) N-HEXADECANE	9.873	7435105	49.731 ug/ml
6) N-OCTADECANE	11.317	7293655	48.412 ug/ml
7) N-EICOSANE	12.628	7010617	47.155 ug/ml
8) N-DOCOSANE	13.829	6680540	45.799 ug/ml
10) N-TETRACOSANE	14.934	6465933	44.824 ug/ml
11) N-HEXACOSANE	15.958	6128537	44.116 ug/ml
12) N-OCTACOSANE	16.910	5938529	44.338 ug/ml
13) N-TRIACONTANE	17.800	5866849	45.370 ug/ml
14) N-DOTRIACONTANE	18.634	5669315	47.119 ug/ml
15) N-TETRATRIACONTANE	19.420	5313983	50.573 ug/ml
16) N-HEXATRIACONTANE	20.162	4768686	49.982 ug/ml
17) N-OCTATRIACONTANE	20.918	4352427	52.667 ug/ml
18) N-TETRACONTANE	21.860	4173084	47.809 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF101920\
Data File : FF007747.D
Signal(s) : FID2B.ch
Acq On : 19 Oct 2020 13:54
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Oct 19 14:07:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF100220.M
Quant Title :
QLast Update : Mon Oct 05 15:41:23 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

