

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF102124\
 Data File : FF014710.D
 Signal(s) : FID2B.ch
 Acq On : 21 Oct 2024 17:15
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Integration File: autoint1.e
 Quant Time: Oct 21 17:35:32 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF102124.M
 Quant Title :
 QLast Update : Mon Oct 21 17:35:19 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.996	691394	5.254 ug/ml
Target Compounds			
1) N-OCTANE	2.044	735139	5.284 ug/ml
2) N-DECANE	4.560	756089	5.293 ug/ml
3) N-DODECANE	6.722	768253	5.294 ug/ml
4) N-TETRADECANE	8.545	775382	5.290 ug/ml
5) N-HEXADECANE	10.150	775794	5.279 ug/ml
6) N-OCTADECANE	11.591	792835	5.281 ug/ml
7) N-EICOSANE	12.900	800046	5.274 ug/ml
8) N-DOCOSANE	14.098	773720	5.258 ug/ml
10) N-TETRACOSANE	15.199	774521	5.270 ug/ml
11) N-HEXACOSANE	16.220	769893	5.271 ug/ml
12) N-OCTACOSANE	17.169	769883	5.328 ug/ml
13) N-TRIACONTANE	18.055	798876	5.457 ug/ml
14) N-DOTRIACONTANE	18.887	801393	5.570 ug/ml
15) N-TETRATRIACONTANE	19.670	752249	5.485 ug/ml
16) N-HEXATRIACONTANE	20.409	661136	5.096 ug/ml
17) N-OCTATRIACONTANE	21.142	614859	5.346 ug/ml
18) N-TETRACONTANE	22.046	557917	5.314 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF102124\
Data File : FF014710.D
Signal(s) : FID2B.ch
Acq On : 21 Oct 2024 17:15
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
5 TRPH STD

Integration File: autoint1.e
Quant Time: Oct 21 17:35:32 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF102124.M
Quant Title :
QLast Update : Mon Oct 21 17:35:19 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

