

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC102218AL\  
 Data File : FC037016.D  
 Signal(s) : FID1A.CH  
 Acq On : 22 Oct 2018 18:49  
 Operator : AJ\SJ  
 Sample : J5605-12  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e  
 Quant Time: Oct 23 04:34:19 2018  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 101718.M  
 Quant Title : GC Extractables  
 QLast Update : Fri Oct 19 15:53:26 2018  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units   |
|--------------------------------|--------|------------|--------------|
| -----                          |        |            |              |
| System Monitoring Compounds    |        |            |              |
| 9) S ortho-Terphenyl (SURR)    | 11.357 | 2239449    | 25.239 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 50.48%       |
| 12) S 1-chlorooctadecane (S... | 12.799 | 1543000    | 24.291 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 48.58%       |
| Target Compounds               |        |            |              |
| 6) T n-Tetradecane (C14)       | 8.088  | 76399      | 1.057 ug/ml  |
| 10) T n-Eicosane (C20)         | 12.394 | 451361     | 6.174 ug/ml  |
| 14) T n-Tetracosane (C24)      | 14.601 | 228198     | 3.219 ug/ml  |
| -----                          |        |            |              |


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC102218AL\  
Data File : FC037016.D  
Signal(s) : FID1A.CH  
Acq On : 22 Oct 2018 18:49  
Operator : AJ\SJ  
Sample : J5605-12  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Integration File: autoint1.e  
Quant Time: Oct 23 04:34:19 2018  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 101718.M  
Quant Title : GC Extractables  
QLast Update : Fri Oct 19 15:53:26 2018  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped


Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#6 n-Tetradecane (C14)

R.T.: 8.088 min  
Delta R.T.: 0.052 min  
Response: 76399  
Conc: 1.06 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 11.357 min  
Delta R.T.: -0.007 min  
Response: 2239449  
Conc: 25.24 ug/ml




#10 n-Eicosane (C20)

R.T.: 12.394 min  
Delta R.T.: 0.029 min  
Response: 451361  
Conc: 6.17 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.799 min  
Delta R.T.: -0.007 min  
Response: 1543000  
Conc: 24.29 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.601 min  
Delta R.T.: -0.056 min  
Response: 228198  
Conc: 3.22 ug/ml