

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102924\
 Data File : FB031056.D
 Signal(s) : FID2B.CH
 Acq On : 29 Oct 2024 10:51
 Operator : YP/AJ
 Sample : VBF1029S2 50X
 Misc : 5.00G/5.00 ML MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :

Integration File: SAMPLE.e
 Quant Time: Oct 29 11:00:10 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102824.M
 Quant Title :
 QLast Update : Mon Oct 28 12:19:57 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.792	559575	20.272 ng/ml
Target Compounds			
7) t Ethylbenzene	13.048	5329	0.119 ng/ml
8) t m-Xylene	13.192	16024	0.329 ng/ml
9) t O-Xylene	13.916	10423	0.229 ng/ml
10) t 1,2,4-Trimethylbenzene	16.204	5967	0.160 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB102924\
Data File : FB031056.D
Signal(s) : FID2B.CH
Acq On : 29 Oct 2024 10:51
Operator : YP/AJ
Sample : VBF1029S2 50X
Misc : 5.00G/5.00 ML MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :

Integration File: SAMPLE.e
Quant Time: Oct 29 11:00:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102824.M
Quant Title :
QLast Update : Mon Oct 28 12:19:57 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT


R.T.: 8.792 min
Delta R.T.: 0.001 min
Response: 559575
Conc: 20.27 ng/ml

Instrument :
FID_F
ClientSampleId :




#7 Ethylbenzene

R.T.: 13.048 min
Delta R.T.: -0.007 min
Response: 5329
Conc: 0.12 ng/ml




#8 m-Xylene

R.T.: 13.192 min
Delta R.T.: 0.004 min
Response: 16024
Conc: 0.33 ng/ml



#9 O-Xylene

R.T.: 13.916 min
Delta R.T.: 0.000 min
Response: 10423
Conc: 0.23 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.204 min
Delta R.T.: 0.012 min
Response: 5967
Conc: 0.16 ng/ml

Instrument :
FID_F
ClientSampleId :