

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF110118\
 Data File : FF002542.D
 Signal(s) : FID2B.ch
 Acq On : 01 Nov 2018 14:37
 Operator : AJ\SJ
 Sample : PB114378BS
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Nov 02 00:54:53 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF101218.M
 Quant Title :
 QLast Update : Fri Oct 12 13:21:40 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.648	2162715	15.560 ug/ml
Target Compounds			
2) N-DECANE	4.375	2800990	18.336 ug/ml
3) N-DODECANE	6.396	2885464	18.634 ug/ml
4) N-TETRADECANE	8.190	2877933	18.391 ug/ml
5) N-HEXADECANE	9.792	2856419	18.174 ug/ml
6) N-OCTADECANE	11.235	2832347	17.822 ug/ml
7) N-EICOSANE	12.546	2761618	17.436 ug/ml
8) N-DOCOSANE	13.747	2685353	17.021 ug/ml
10) N-TETRACOSANE	14.853	2613121	16.702 ug/ml
11) N-HEXACOSANE	15.876	2517280	16.166 ug/ml
12) N-OCTACOSANE	16.829	2460256	15.861 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : \\74.0.250.170\Terastorage\pestpcbsrv\HPCHEM1\FID_F\Data\FF110118\
Data File : FF002542.D
Signal(s) : FID2B.ch
Acq On : 01 Nov 2018 14:37
Operator : AJ\SJ
Sample : PB114378BS
Misc :
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Nov 02 00:54:53 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF101218.M
Quant Title :
QLast Update : Fri Oct 12 13:21:40 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

